

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.052 °C	30.4 °C	0.348 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.691 %HR	27 %HR	-2.691 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.499 %HR	47 %HR	-3.499 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.223 %HR	70 %HR	-0.223 %HR	± 5.0 %HR	± 1.9E+00 %HR