

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
49,488.8 lx	49,500 lx	-11.2 lx	± 1982 lx	±	4.2E+02 lx
10,510.0 lx	10,520 lx	-10.0 lx	± 423 lx	±	3.7E+01 lx
5,646.6 lx	5,654 lx	-7.4 lx	± 228 lx	±	1.5E+01 lx
1,035.5 lx	1,040.0 lx	-4.5 lx	± 44 lx	±	6.1E-01 lx
340.7 lx	343.3 lx	-2.6 lx	± 16 lx	±	8.8E-02 lx