

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.75	0.750011	0.150	± 0.9	±	4.3E-01 mm						
0 a 0.8	0.800021	0.160	± 0.9	±	4.3E-01 mm						
0 a 0.9	0.900028	0.200	± 0.9	±	4.3E-01 mm						
0 a 1	1.000036	0.250	± 1	±	4.3E-01 mm						
0 a 1.2	1.200055	0.350	± 1	±	4.8E-01 mm						
0 a 1.4	1.400080	0.470	± 1.1	±	4.8E-01 mm						
0 a 1.6	1.600087	0.620	± 1.2	±	4.8E-01 mm						
0 a 1.85	1.850045	0.900	± 1.3	±	4.8E-01 mm						
0 a 2	2.000058	1.040	± 1.4	±	4.8E-01 mm						