

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.75	0.750004	0.150	± 0.9	±	4.3E-01 mm						
0 a 0.8	0.800055	0.120	± 0.9	±	4.3E-01 mm						
0 a 0.9	0.900064	0.160	± 0.9	±	4.3E-01 mm						
0 a 1	1.000086	0.200	± 1	±	4.3E-01 mm						
0 a 1.2	1.200099	0.310	± 1	±	4.8E-01 mm						
0 a 1.4	1.400184	0.360	± 1.1	±	4.8E-01 mm						
0 a 1.6	1.600017	0.690	± 1.2	±	4.8E-01 mm						
0 a 1.85	1.850111	0.830	± 1.3	±	4.8E-01 mm						
0 a 2	2.000079	1.020	± 1.4	±	4.8E-01 mm						