

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.95 ml	100 ml	0.053 ml	± 1.0 ml	±	1.4E-01 µl
500.23 ml	500 ml	-0.234 ml	± 5.0 ml	±	1.4E-01 µl
1,001.20 ml	1000 ml	-1.200 ml	± 10.0 ml	±	1.5E-01 µl