

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1 mm	0.99 mm	-10 µm	± 50 µm	±	5.8E+00 µm
2 mm	1.99 mm	-10 µm	± 50 µm	±	5.8E+00 µm
3 mm	2.99 mm	-10 µm	± 50 µm	±	5.8E+00 µm
5 mm	5.00 mm	0 µm	± 50 µm	±	5.8E+00 µm
10 mm	10.00 mm	0 µm	± 50 µm	±	5.8E+00 µm
15 mm	15.00 mm	0 µm	± 50 µm	±	5.8E+00 µm
20 mm	20.00 mm	0 µm	± 50 µm	±	5.8E+00 µm
25 mm	25.00 mm	0 µm	± 50 µm	±	5.8E+00 µm
50.00 mm	49.98 mm	-20 µm	± 50 µm	±	5.8E+00 µm