

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.671 °C	29.8 °C	0.129 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.756 %HR	33.7 %HR	1.944 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.110 %HR	50.9 %HR	0.790 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.460 %HR	77.3 %HR	0.840 %HR	± 3.0 %HR	± 1.9E+00 %HR