

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.75	0.750001	0.270	± 0.3	±	4.3E-01 mm						
0 a 0.8	0.800004	0.270	± 0.3	±	4.3E-01 mm						
0 a 0.9	0.900005	0.270	± 0.3	±	4.3E-01 mm						
0 a 1	1.000004	0.280	± 1.0	±	4.3E-01 mm						
0 a 1.2	1.200003	0.290	± 1.0	±	4.8E-01 mm						
0 a 1.4	1.400006	0.290	± 0.4	±	4.8E-01 mm						
0 a 1.6	1.600006	0.300	± 0.5	±	4.8E-01 mm						
0 a 1.85	1.850000	0.320	± 0.5	±	4.8E-01 mm						
0 a 2	2.000002	0.330	± 0.6	±	4.8E-01 mm						