

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.671 °C	29.7 °C	0.029 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.756 %HR	33.9 %HR	2.144 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.110 %HR	48.3 %HR	-1.810 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.460 %HR	74.0 %HR	-2.460 %HR	± 3.0 %HR	± 1.9E+00 %HR