

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.850 °C	14.7 °C	-0.150 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.721 °C	30.0 °C	0.279 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.527 %HR	52.2 %HR	2.673 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.662 %HR	68.5 %HR	-2.162 %HR	± 3.0 %HR	± 1.9E+00 %HR