

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.850 °C	15.0 °C	0.150 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.721 °C	29.8 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.527 %HR	51.9 %HR	2.373 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.662 %HR	71.5 %HR	0.838 %HR	± 3.0 %HR	± 1.9E+00 %HR