

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.1	0.09980	0.200	± 0.2	±	4.3E-01 mm						
0 a 0.5	0.49979	0.210	± 0.3	±	4.3E-01 mm						
0 a 1	0.99973	0.270	± 0.4	±	4.3E-01 mm						
0 a 2	1.99963	0.370	± 0.6	±	4.8E-01 mm						
0 a 3	2.99928	0.720	± 0.8	±	4.8E-01 mm						
0 a 4	3.99928	0.720	± 1	±	4.8E-01 mm						
0 a 5	4.99930	0.700	± 1.2	±	4.8E-01 mm						
0 a 6	5.99937	0.630	± 1.4	±	3.6E-01 mm						
0 a 7	6.99939	0.610	± 1.6	±	3.6E-01 mm						
0 a 8	7.99937	0.630	± 1.8	±	3.6E-01 mm						