

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.641 °C	-25.9 °C	-0.259 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.020 °C	-0.2 °C	-0.180 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.943 °C	4.9 °C	-0.043 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.957 °C	10.0 °C	0.043 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.1 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.040 °C	40.4 °C	0.360 °C	± 0.5 °C	± 4.0E-01 °C

