

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.641 °C	-25.5 °C	0.141 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	-0.020 °C	0.0 °C	0.020 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.943 °C	5.0 °C	0.057 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.957 °C	10.0 °C	0.043 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.0 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.3 °C	0.238 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.040 °C	40.3 °C	0.260 °C	± 0.5 °C	± 4.0E-01 °C

