

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
53.00 mmHg	50 mmHg	-3.00 mmHg	± 3 mmHg	±	6.6E-01 mmHg
104.00 mmHg	100 mmHg	-4.00 mmHg	± 3 mmHg	±	6.5E-01 mmHg
153.00 mmHg	150 mmHg	-3.00 mmHg	± 3 mmHg	±	6.6E-01 mmHg
204.00 mmHg	200 mmHg	-4.00 mmHg	± 3 mmHg	±	6.8E-01 mmHg
253.00 mmHg	250 mmHg	-3.00 mmHg	± 3 mmHg	±	7.5E-01 mmHg
299.00 mmHg	300 mmHg	1.00 mmHg	± 3 mmHg	±	7.5E-01 mmHg