

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.721 °C	29.6 °C	-0.121 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	32.124 %HR	27 %HR	-5.124 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.527 %HR	44 %HR	-5.527 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.685 %HR	65 %HR	-6.685 %HR	± 5.0 %HR	± 1.9E+00 %HR