

Certificado No. (Certificate #):63731Equipo (Instrument):Medidor de Espesores

Fecha de Calibración (Calibration Date):2025-12-02Marca (Brand):YOWEXA

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50 μm	50.0 μm	0.0 μm	±(2%+1μm)	±	1.1E+00 μm
100 μm	100 μm	0 μm	±(2%+1μm)	±	1.3E+00 μm
150 μm	150 μm	-1 μm	±(2%+1μm)	±	1.1E+00 μm
248 μm	248 μm	0 μm	±(2%+1μm)	±	1.6E+00 μm
398 μm	397 μm	-1 μm	±(2%+1μm)	±	1.6E+00 μm
496 μm	496 μm	0 μm	±(2%+1μm)	±	1.7E+00 μm
991 μm	993 μm	2 μm	±(2%+1μm)	±	1.9E+00 μm
1,389 μm	1,388 μm	-1 μm	±(2%+1μm)	±	1.7E+00 μm
1,885 μm	1,886 μm	1 μm	±(2%+1μm)	±	2.1E+00 μm