

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-10 °C	-10.048 °C	-10.1 °C	-0.052 °C	± 2.0 °C	± 4.0E-01 °C
0 °C	-0.020 °C	0.1 °C	0.120 °C	± 2.0 °C	± 4.0E-01 °C
2 °C	2.486 °C	2.6 °C	0.114 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.1 °C	0.093 °C	± 2.0 °C	± 4.0E-01 °C
15 °C	15.422 °C	15.3 °C	-0.122 °C	± 2.0 °C	± 4.0E-01 °C
40 °C	40.591 °C	40.3 °C	-0.291 °C	± 2.0 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	21.6 %HR	1.255 %HR	± 10.0 %HR	± 1.9E+00 %HR
50 %HR	50.407 %HR	47.9 %HR	-2.507 %HR	± 10.0 %HR	± 1.9E+00 %HR
95 %HR	95.120 %HR	92.8 %HR	-2.320 %HR	± 10.0 %HR	± 1.9E+00 %HR