

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.792 °C	29.8 °C	0.008 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.480 %HR	28.8 %HR	-2.680 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.251 %HR	51.8 %HR	2.549 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.272 %HR	77.2 %HR	2.928 %HR	± 3.0 %HR	± 1.9E+00 %HR