

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.792 °C	29.8 °C	0.008 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.480 %HR	28.7 %HR	-2.780 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.251 %HR	51.8 %HR	2.549 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.272 %HR	77.0 %HR	2.728 %HR	± 3.0 %HR	± 1.9E+00 %HR