

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.792 °C	29.7 °C	-0.092 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.480 %HR	29.2 %HR	-2.280 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.251 %HR	52.0 %HR	2.749 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.272 %HR	77.2 %HR	2.928 %HR	± 3.0 %HR	± 1.9E+00 %HR