

<b>Patrón (Standard)</b>	<b>UBP (UIT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
0.4908 ml	0.5000 ml	0.0092 ml	± 0.03 ml	± 5.3E-02 µl
2.4903 ml	2.5000 ml	0.0097 ml	± 0.03 ml	± 1.1E-01 µl
5.0006 ml	5.0000 ml	-0.0006 ml	± 0.03 ml	± 1.7E-01 µl