

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
2.4960 ml	2.5000 ml	0.0040 ml	± 0.2 ml	±	1.1E-01 µl
12.4944 ml	12.5000 ml	0.0056 ml	± 0.2 ml	±	3.9E-01 µl
24.9110 ml	25.0000 ml	0.0890 ml	± 0.2 ml	±	7.7E-01 µl