

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.0081 ml	10.0000 ml	-0.0081 ml	± 0.1 ml	±	3.1E-01 µl
50.0068 ml	50.0000 ml	-0.0068 ml	± 0.1 ml	±	1.5E+00 µl
100.0525 ml	100.0000 ml	-0.0525 ml	± 0.1 ml	±	3.1E+00 µl