

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.952 °C	29.9 °C	-0.052 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	27.3 %HR	-2.237 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	50.928 %HR	48.5 %HR	-2.428 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	70.335 %HR	67.9 %HR	-2.435 %HR	± 2.5 %HR	± 1.9E+00 %HR