

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.931 °C	14.9 °C	-0.031 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.924 °C	19.9 °C	-0.024 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	27.3 %HR	-2.237 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	50.928 %HR	48.5 %HR	-2.428 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	70.335 %HR	67.90 %HR	-2.435 %HR	± 2.5 %HR	± 1.9E+00 %HR