

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.931 °C	15.3 °C	0.369 °C	± 0.8 °C	± 4.0E-01 °C
20 °C	19.924 °C	20.3 °C	0.376 °C	± 0.8 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	26.1 %HR	-3.437 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	50.928 %HR	47.5 %HR	-3.428 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.335 %HR	66.8 %HR	-3.535 %HR	± 4.0 %HR	± 1.9E+00 %HR