

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.03	0.02988	0.120	± 0.2	±	4.3E-01 mm						
0 a 0.06	0.05981	0.190	± 0.2	±	4.3E-01 mm						
0 a 0.09	0.08981	0.190	± 0.2	±	4.3E-01 mm						
0 a 0.12	0.11980	0.200	± 0.2	±	4.3E-01 mm						
0 a 0.15	0.14980	0.200	± 0.2	±	4.3E-01 mm						
0 a 0.18	0.17980	0.200	± 0.2	±	4.3E-01 mm						
0 a 0.21	0.20980	0.200	± 0.2	±	4.3E-01 mm						
0 a 0.24	0.23981	0.190	± 0.2	±	4.3E-01 mm						
0 a 0.27	0.26981	0.190	± 0.2	±	4.3E-01 mm						
0 a 0.3	0.29980	0.200	± 0.2	±	4.3E-01 mm						