

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	0.99938	0.620	± 1	± 4.3E-01 mm					
0 a 2	1.99978	0.220	± 1.4	± 4.8E-01 mm					
0 a 3	2.99951	0.490	± 1.8	± 4.8E-01 mm					
0 a 4	3.99987	0.130	± 2.2	± 4.8E-01 mm					
0 a 5	4.99976	0.240	± 2.6	± 4.8E-01 mm					
0 a 6	5.99990	0.100	± 3	± 3.6E-01 mm					
0 a 7	6.99989	0.110	± 3.4	± 3.6E-01 mm					
0 a 8	8.00017	-0.170	± 3.8	± 3.6E-01 mm					
0 a 9	8.99989	0.110	± 4.2	± 3.6E-01 mm					
0 a 10	10.00015	-0.150	± 4.6	± 3.6E-01 mm					
0 a 11	10.99988	0.120	± 5	± 3.6E-01 mm					
0 a 12	12.00022	-0.220	± 5.4	± 3.6E-01 mm					
0 a 13	13.00020	-0.200	± 5.8	± 3.6E-01 mm					
0 a 14	14.00051	-0.510	± 6.2	± 3.6E-01 mm					
0 a 15	15.00028	-0.280	± 6.6	± 3.6E-01 mm					
0 a 16	16.00034	-0.340	± 7	± 3.6E-01 mm					
0 a 17	17.00006	-0.060	± 7.4	± 3.6E-01 mm					
0 a 18	18.00013	-0.130	± 7.8	± 3.6E-01 mm					
0 a 19	19.00031	-0.310	± 8.2	± 3.6E-01 mm					
0 a 20	20.00093	-0.930	± 8.6	± 3.6E-01 mm					
0 a 21	21.00064	-0.640	± 9	± 5.5E-01 mm					
0 a 22	22.00062	-0.620	± 9.4	± 5.5E-01 mm					
0 a 23	23.00050	-0.500	± 9.8	± 5.5E-01 mm					
0 a 24	24.00066	-0.660	± 10.2	± 5.5E-01 mm					
0 a 25	25.00058	-0.580	± 10.6	± 5.5E-01 mm					
0 a 26	26.00087	-0.870	± 11	± 5.5E-01 mm					
0 a 27	27.00068	-0.680	± 11.4	± 5.5E-01 mm					
0 a 28	28.00114	-1.140	± 11.8	± 5.5E-01 mm					
0 a 29	29.00092	-0.920	± 12.2	± 5.5E-01 mm					
0 a 30	30.00133	-1.330	± 12.6	± 5.5E-01 mm					