

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49921	0.790	± 0.8	±	4.3E-01 mm						
0 a 1	0.99920	0.800	± 1	±	4.3E-01 mm						
0 a 1.5	1.49908	0.920	± 1.2	±	4.8E-01 mm						
0 a 2	1.99912	0.880	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49919	0.810	± 1.6	±	4.8E-01 mm						
0 a 3	2.99905	0.950	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49933	0.670	± 2	±	4.8E-01 mm						
0 a 4	3.99912	0.880	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49930	0.700	± 2.4	±	4.8E-01 mm						
0 a 5	4.99919	0.810	± 2.6	±	4.8E-01 mm						