

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 1 mm                     | 1.00 mm          | 0 µm                 | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 1.25 mm                  | 1.25 mm          | 0 µm                 | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 1.50 mm                  | 1.51 mm          | 10 µm                | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 2 mm                     | 2.00 mm          | 0 µm                 | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 3 mm                     | 3.00 mm          | 0 µm                 | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 5 mm                     | 5.01 mm          | 10 µm                | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 8 mm                     | 8.01 mm          | 10 µm                | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 10 mm                    | 10.01 mm         | 10 µm                | ± 20 µm                     | ±                                  | 5.8E+00 µm |
| 12 mm                    | 12.01 mm         | 10 µm                | ± 20 µm                     | ±                                  | 5.8E+00 µm |