

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
5.00 mm	5.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
10.00 mm	10.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
15.00 mm	15.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
20.00 mm	20.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
25.00 mm	25.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
30.00 mm	30.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
40.00 mm	40.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
50.00 mm	50.02 mm	20 µm	± 30 µm	±	5.8E+00 µm