

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.0 μm	11 μm	1.0 μm	± 0.55 μm	±	5.8E-01 μm
12.0 μm	13 μm	1.0 μm	± 0.65 μm	±	5.8E-01 μm
20.0 μm	22 μm	2.0 μm	± 1.1 μm	±	5.8E-01 μm