

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.764 °C	1.5 °C	-0.264 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.771 °C	7.5 °C	-0.271 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.027 °C	10.0 °C	-0.027 °C	± 0.5 °C	± 4.0E-01 °C