

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.264 °C	-31.0 °C	-0.736 °C	± 0.6 °C	± 4.0E-01 °C
-25 °C	-25.270 °C	-26.0 °C	-0.730 °C	± 0.6 °C	± 4.0E-01 °C
-15 °C	-15.343 °C	-16.0 °C	-0.657 °C	± 0.6 °C	± 4.0E-01 °C
-10 °C	-10.289 °C	-11.0 °C	-0.711 °C	± 0.6 °C	± 4.0E-01 °C
2 °C	1.754 °C	1.0 °C	-0.754 °C	± 0.6 °C	± 4.0E-01 °C
8 °C	7.781 °C	7.0 °C	-0.781 °C	± 0.6 °C	± 4.0E-01 °C