

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.264 °C	-30.5 °C	-0.236 °C	± 0.6 °C	± 4.0E-01 °C
-25 °C	-25.270 °C	-25.5 °C	-0.230 °C	± 0.6 °C	± 4.0E-01 °C
-15 °C	-15.343 °C	-15.5 °C	-0.157 °C	± 0.6 °C	± 4.0E-01 °C
-10 °C	-10.289 °C	-10.5 °C	-0.211 °C	± 0.6 °C	± 4.0E-01 °C
2 °C	1.754 °C	1.5 °C	-0.254 °C	± 0.6 °C	± 4.0E-01 °C
8 °C	7.781 °C	7.5 °C	-0.281 °C	± 0.6 °C	± 4.0E-01 °C