

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
2 °C	1.764 °C	2.0 °C	0.236 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.771 °C	8.0 °C	0.229 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.027 °C	10.5 °C	0.473 °C	± 0.5 °C	± 4.0E-01 °C