

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.264 °C	-30.3 °C	-0.036 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.270 °C	-25.3 °C	-0.030 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.343 °C	-15.5 °C	-0.157 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.289 °C	-10.1 °C	0.189 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.754 °C	1.7 °C	-0.054 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.781 °C	7.6 °C	-0.181 °C	± 0.5 °C	± 4.0E-01 °C