

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.312 °C	15.4 °C	0.088 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.129 °C	25.4 °C	0.271 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.812 °C	30.3 °C	0.488 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.282 %HR	57.2 %HR	-2.082 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.425 %HR	63.6 %HR	-0.825 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.744 %HR	69.6 %HR	-1.144 %HR	± 3.0 %HR	± 1.9E+00 %HR