

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.312 °C	15.2 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.129 °C	25.3 °C	0.171 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.812 °C	30.1 °C	0.288 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.282 %HR	56.4 %HR	-2.882 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.425 %HR	60.2 %HR	-4.225 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.744 %HR	64.8 %HR	-5.944 %HR	± 3.0 %HR	± 1.9E+00 %HR