

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-10 °C	-10.048 °C	-10.3 °C	-0.252 °C	± 2.0 °C	± 4.0E-01 °C
2 °C	2.486 °C	2.4 °C	-0.086 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.0 °C	-0.007 °C	± 2.0 °C	± 4.0E-01 °C