

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -10 °C            | -10.048 °C        | -10.2 °C  | -0.152 °C     | ± 2.0 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.020 °C         | 0.0 °C    | 0.020 °C      | ± 2.0 °C             | ± 4.0E-01 °C                |
| 2 °C              | 2.486 °C          | 2.6 °C    | 0.114 °C      | ± 2.0 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.007 °C         | 9.9 °C    | -0.107 °C     | ± 2.0 °C             | ± 4.0E-01 °C                |
| 15 °C             | 15.422 °C         | 15.4 °C   | -0.022 °C     | ± 2.0 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.591 °C         | 40.0 °C   | -0.591 °C     | ± 2.0 °C             | ± 4.0E-01 °C                |
| 20 %HR            | 20.345 %HR        | 21.9 %HR  | 1.555 %HR     | ± 10.0 %HR           | ± 1.9E+00 %HR               |
| 50 %HR            | 50.407 %HR        | 48.4 %HR  | -2.007 %HR    | ± 10.0 %HR           | ± 1.9E+00 %HR               |
| 95 %HR            | 95.120 %HR        | 92.7 %HR  | -2.420 %HR    | ± 10.0 %HR           | ± 1.9E+00 %HR               |