

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-10 °C	-10.048 °C	-10.2 °C	-0.152 °C	± 2.0 °C	± 4.0E-01 °C
0 °C	-0.020 °C	0.0 °C	0.020 °C	± 2.0 °C	± 4.0E-01 °C
2 °C	2.486 °C	2.3 °C	-0.186 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.9 °C	-0.107 °C	± 2.0 °C	± 4.0E-01 °C
15 °C	15.422 °C	15.4 °C	-0.022 °C	± 2.0 °C	± 4.0E-01 °C
40 °C	40.591 °C	40.4 °C	-0.191 °C	± 2.0 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	19.6 %HR	-0.745 %HR	± 10.0 %HR	± 1.9E+00 %HR
50 %HR	50.407 %HR	47.2 %HR	-3.207 %HR	± 10.0 %HR	± 1.9E+00 %HR
95 %HR	95.120 %HR	92.2 %HR	-2.920 %HR	± 10.0 %HR	± 1.9E+00 %HR