

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-10 °C	-10.048 °C	-10.3 °C	-0.252 °C	± 2.0 °C	± 4.0E-01 °C
0 °C	-0.020 °C	0.0 °C	0.020 °C	± 2.0 °C	± 4.0E-01 °C
2 °C	2.486 °C	2.2 °C	-0.286 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.8 °C	-0.207 °C	± 2.0 °C	± 4.0E-01 °C
15 °C	15.422 °C	15.5 °C	0.078 °C	± 2.0 °C	± 4.0E-01 °C
40 °C	40.591 °C	40.4 °C	-0.191 °C	± 2.0 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	20.9 %HR	0.555 %HR	± 10.0 %HR	± 1.9E+00 %HR
50 %HR	50.407 %HR	48.8 %HR	-1.607 %HR	± 10.0 %HR	± 1.9E+00 %HR
95 %HR	95.120 %HR	92.7 %HR	-2.420 %HR	± 10.0 %HR	± 1.9E+00 %HR