

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
12.2 RPM	12 RPM	-0.2 RPM	± 1 RPM	±	1.6E-01 RPM
29.92 RPM	30 RPM	0.08 RPM	± 1 RPM	±	1.2E-01 RPM
60.04 RPM	60 RPM	-0.04 RPM	± 1 RPM	±	1.1E-01 RPM