

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.842 °C	30.2 °C	0.358 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	27 %HR	-2.537 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.928 %HR	49 %HR	-1.928 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.335 %HR	73 %HR	2.665 %HR	± 5.0 %HR	± 1.9E+00 %HR