

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.842 °C	29 °C	-0.842 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	25 %HR	-4.537 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.928 %HR	46 %HR	-4.928 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.335 %HR	65 %HR	-5.335 %HR	± 5.0 %HR	± 1.9E+00 %HR