

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
200 RPM	200 RPM	0 RPM	± 10 RPM	±	5.8E+01 RPM
599 RPM	600 RPM	1 RPM	± 30 RPM	±	5.8E+01 RPM
999 RPM	1000 RPM	1 RPM	± 50 RPM	±	5.8E+01 RPM
1399 RPM	1400 RPM	1 RPM	± 70 RPM	±	5.8E+01 RPM
1799 RPM	1800 RPM	1 RPM	± 90 RPM	±	5.8E+01 RPM