

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.75	0.750009	0.150	± 0.9	±	4.3E-01 mm						
0 a 0.8	0.800075	0.100	± 0.9	±	4.3E-01 mm						
0 a 0.9	0.900084	0.140	± 0.9	±	4.3E-01 mm						
0 a 1	1.000099	0.180	± 1	±	4.3E-01 mm						
0 a 1.2	1.200108	0.300	± 1	±	4.8E-01 mm						
0 a 1.4	1.400129	0.420	± 1.1	±	4.8E-01 mm						
0 a 1.6	1.600016	0.700	± 1.2	±	4.8E-01 mm						
0 a 1.85	1.850177	0.770	± 1.3	±	4.8E-01 mm						
0 a 2	2.000079	1.020	± 1.4	±	4.8E-01 mm						