

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	3.991 °C	4.5 °C	0.509 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.222 °C	15.5 °C	0.278 °C	± 3.0 °C	± 4.0E-01 °C
30 °C	30.163 °C	29.7 °C	-0.463 °C	± 3.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	27.5 %HR	-2.866 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.100 %HR	47.3 %HR	-2.800 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.294 %HR	62.9 %HR	-2.394 %HR	± 3.0 %HR	± 1.9E+00 %HR